



**Supplemental/Bid Bulletin No. 1**  
**October 23, 2019**

**PUBLIC BIDDING No. 19-304-10**

**Supply and Delivery of Various Medical Equipment for**  
**Margosatubig Regional Hospital (MRH)**

Issued pursuant to Sec. 22.5 of the IRR of R.A. 9184 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by prospective bidder/s during the Pre-Bid Conference held on **October 16, 2019**.

**A. AMENDMENTS/INCLUSIONS**

REFERENCE		BASIS FOR AMENDMENT/INCLUSIONS
<b>Section VII. Technical Specifications</b>		
1.	LOT NO. 2                      Emergency Cart QUANTITY                      1 unit APPROVED BUDGET            P 115,000.00 FOR THE CONTRACT	<i>To amend the Requirement as per End-user's reply dated October 23, 2019</i>
	AGENCY'S SPECIFICATIONS	
	Item Description	
	Dimension (mm): 850x520x1045-950 (min)	
	XXX	
	Guard Rail: Aluminum or Stainless Steel Drawer: ABS Plastic • <del>1 small drawer, 3 medium drawers, 1 large drawer</del> • <u>1-2 small drawer, 2-3 medium drawer, 1 large drawer with a total of 5 drawers</u>	
2.	LOT NO. 2                      Laryngoscope, Pedia- Miller QUANTITY                      3 units APPROVED BUDGET            P 130,000.00 FOR THE CONTRACT	<i>To amend the Requirement as per End-user's reply dated October 23, 2019</i>
	AGENCY'S SPECIFICATIONS	
	XXX	
	Standard Requirement	

	xxx <del>1. One (1) year warranty on parts and services for the Laryngoscopes. Warranty period shall commence from the date of acceptance by the end user after testing.</del>															
3	<table border="0"> <tr> <td>LOT NO. 7</td> <td>Electrocardiograph Machine with Analyzer</td> </tr> <tr> <td>QUANTITY</td> <td>2 units</td> </tr> <tr> <td>APPROVED BUDGET FOR THE CONTRACT</td> <td>P 346,153.00</td> </tr> </table> <table border="1"> <thead> <tr> <th colspan="2">AGENCY'S SPECIFICATIONS</th> </tr> </thead> <tbody> <tr> <td>Item Description</td> <td></td> </tr> <tr> <td>xxx</td> <td></td> </tr> <tr> <td>Sensitivity of Gain</td> <td>x 1/4 to x 2</td> </tr> </tbody> </table>	LOT NO. 7	Electrocardiograph Machine with Analyzer	QUANTITY	2 units	APPROVED BUDGET FOR THE CONTRACT	P 346,153.00	AGENCY'S SPECIFICATIONS		Item Description		xxx		Sensitivity of Gain	x 1/4 to x 2	To amend the Requirement as per End-user's reply dated October 23, 2019
LOT NO. 7	Electrocardiograph Machine with Analyzer															
QUANTITY	2 units															
APPROVED BUDGET FOR THE CONTRACT	P 346,153.00															
AGENCY'S SPECIFICATIONS																
Item Description																
xxx																
Sensitivity of Gain	x 1/4 to x 2															
4	<table border="0"> <tr> <td>LOT NO. 9</td> <td>EENT Diagnostic Set</td> </tr> <tr> <td>QUANTITY</td> <td>7 units-sets</td> </tr> <tr> <td>APPROVED BUDGET FOR THE CONTRACT</td> <td>P 350,000.00</td> </tr> </table>	LOT NO. 9	EENT Diagnostic Set	QUANTITY	7 units-sets	APPROVED BUDGET FOR THE CONTRACT	P 350,000.00	To amend the Requirement as per End-user's reply dated October 23, 2019								
LOT NO. 9	EENT Diagnostic Set															
QUANTITY	7 units-sets															
APPROVED BUDGET FOR THE CONTRACT	P 350,000.00															

**B. DURING PRE-BID CONFERENCE**

DETAIL		CLARIFICATION				
1	<p>The Representative from Actimed requested to modify the below requirement:</p> <table border="1"> <thead> <tr> <th>From</th> <th>To</th> </tr> </thead> <tbody> <tr> <td>Lot No. 2 Dimension (mm): 850 x 520 x 1045</td> <td>Lot No. 2 Dimension (mm): 850 x 520 x 950 (min)</td> </tr> </tbody> </table>	From	To	Lot No. 2 Dimension (mm): 850 x 520 x 1045	Lot No. 2 Dimension (mm): 850 x 520 x 950 (min)	Please refer to the above amendment.
From	To					
Lot No. 2 Dimension (mm): 850 x 520 x 1045	Lot No. 2 Dimension (mm): 850 x 520 x 950 (min)					
2	<p>The Representative from Actimed requested to modify the below requirement:</p> <table border="1"> <thead> <tr> <th>From</th> <th>To</th> </tr> </thead> <tbody> <tr> <td>Lot No. 2 Guard rail: Aluminum</td> <td>Lot No. 2 Guard Rail: Aluminum or Stainless Steel</td> </tr> </tbody> </table>	From	To	Lot No. 2 Guard rail: Aluminum	Lot No. 2 Guard Rail: Aluminum or Stainless Steel	Please refer to the above amendment.
From	To					
Lot No. 2 Guard rail: Aluminum	Lot No. 2 Guard Rail: Aluminum or Stainless Steel					

3	<p>The Representative from Actimed requested to modify the below requirement:</p> <table border="1" data-bbox="294 443 920 678"> <thead> <tr> <th data-bbox="294 443 609 479">From</th> <th data-bbox="609 443 920 479">To</th> </tr> </thead> <tbody> <tr> <td data-bbox="294 479 609 678">           Lot No. 2            Drawer: ABS Plastic            1 small drawer, 3            medium drawers, 1            large drawer         </td> <td data-bbox="609 479 920 678">           Lot No. 2            Drawer: ABS Plastic            2 small drawers, 2            medium drawers, 1            large drawer         </td> </tr> </tbody> </table>	From	To	Lot No. 2 Drawer: ABS Plastic 1 small drawer, 3 medium drawers, 1 large drawer	Lot No. 2 Drawer: ABS Plastic 2 small drawers, 2 medium drawers, 1 large drawer	<p><i>Please refer to the above amendment.</i></p>
From	To					
Lot No. 2 Drawer: ABS Plastic 1 small drawer, 3 medium drawers, 1 large drawer	Lot No. 2 Drawer: ABS Plastic 2 small drawers, 2 medium drawers, 1 large drawer					
4	<p>The Representative from Actimed requested to modify the below requirement:</p> <table border="1" data-bbox="294 846 920 1014"> <thead> <tr> <th data-bbox="294 846 609 882">From</th> <th data-bbox="609 846 920 882">To</th> </tr> </thead> <tbody> <tr> <td data-bbox="294 882 609 1014">           Lot No. 5            Bolus Rate: 1-            999.9ml/h, 1-999.9ml         </td> <td data-bbox="609 882 920 1014">           Lot No. 5            Bolus Rate: 100-            1500ml         </td> </tr> </tbody> </table>	From	To	Lot No. 5 Bolus Rate: 1- 999.9ml/h, 1-999.9ml	Lot No. 5 Bolus Rate: 100- 1500ml	<p><i>The Original Technical Specification is retained as per End-user's reply dated October 23, 2019.</i></p>
From	To					
Lot No. 5 Bolus Rate: 1- 999.9ml/h, 1-999.9ml	Lot No. 5 Bolus Rate: 100- 1500ml					
5	<p>All Bidders:</p> <p>Clarification on the unit of measurement of the below requirement:</p> <p>Lot No. 7            Sensitivity x1/4 to x2</p>	<p><i>Please refer to the above amendment.</i></p>				

**C. CLARIFICATIONS**

DETAIL	CLARIFICATION
<p><b>Letter from <sup>1</sup> MACTYCOON MARKETING</b></p>	
<p>1 For Lot. 4, 2 units Ventilator – the ABC is not enough for a 2 unit ventilator, this is probably just for 1 unit.</p>	<p><i>Changes in the ABC or quantity is not applicable as of this stage of the bidding process</i></p>

<sup>1</sup> Email letter dated October 21, 2019

2	For all the Lots, is the calibration and calibration certificate included in the preventive maintenance requirement of the machine, and if the calibration is included in the preventive maintenance, would you be requiring a Biomedical Engineer with calibration license to be employed by the bidder.	<b><i>Issuance of a "Factory" Calibration Certificate is only required during the delivery of the equipment, while the preventive maintenance is scheduled quarterly and a Biomedical Engineer with calibration license is not required.</i></b>
---	---	--

By Rule, all other related provisions in the Bidding Documents correspondingly affected by these amendments are likewise deemed amended to conform to this Bid Bulletin.

Amendments made herein shall be considered an integral part of the Bidding Documents.

**SGD**  
**ENGR. PABLO ROMAN C. ANDRES**  
*Chairperson, PS BACX*

*For the purpose of this Bulletin and for better understanding of its contents, the following rules shall apply: (a) ~~Double Strike-out~~ — denotes deletion; (b) Underline — denotes inclusion or new item/requirement; and "xxx" — denotes separation of phrase/s being amended from the rest of the main text.*