



Republic of the Philippines
 Department of Budget and Management
PROCUREMENT SERVICE
BIDS AND AWARDS COMMITTEE



Bid Bulletin No. 1
 November 6, 2018

PUBLIC BIDDING NO. 18-350-5

**SUPPLY AND DELIVERY OF BLOOD BANK REFRIGERATOR FOR THE
 DEPARTMENT OF HEALTH – NATIONAL VOLUNTARY BLOOD SERVICES
 PROGRAM**

Issued pursuant to Sec. 22.5 of the IRR of R.A. 9184 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by prospective bidders during the **Pre-Bid Conference** held on **24 October 2018**, likewise, respond to bidders' written queries received within the prescriptive period for filing.

A. AMENDMENTS

REFERENCE		BASES FOR AMENDMENT											
SECTION III. BID DATA SHEET													
BDS Clause 5.4 page 40 The Bidder must have completed, within the period specified in the Invitation to Bid and ITB Clause 12.1(a)(ii), a single contract that is similar to this Project, equivalent to at least twenty-five percent (25%) fifty percent (50%) of the ABC.		To amend the requirement as clarified by the end-user											
SECTION V. SPECIAL CONDITIONS OF CONTRACT													
SCC Clause 1.1(k) Page 64-66 The Project Sites: <table border="1" data-bbox="220 1686 1058 2040"> <thead> <tr> <th>Lot No.</th> <th>Quantity</th> <th>Blood Service Facility</th> <th>Contact Person</th> <th>Address</th> </tr> </thead> <tbody> <tr> <td>1</td> <td>1 unit</td> <td>Lupon District Hospital Davao Oriental Provincial Hospital - Lupon (former Lupon District Hospital)</td> <td>Dr. Raul Failma (087) 808-0122 Dr. Emmanuel Pinsoy (087) 808-0122/0920-765-0835</td> <td>Townsite, Lupon, Davao Oriental</td> </tr> </tbody> </table>		Lot No.	Quantity	Blood Service Facility	Contact Person	Address	1	1 unit	Lupon District Hospital Davao Oriental Provincial Hospital - Lupon (former Lupon District Hospital)	Dr. Raul Failma (087) 808-0122 Dr. Emmanuel Pinsoy (087) 808-0122/0920-765-0835	Townsite, Lupon, Davao Oriental	To clarify the requirement for better understanding.	
Lot No.	Quantity	Blood Service Facility	Contact Person	Address									
1	1 unit	Lupon District Hospital Davao Oriental Provincial Hospital - Lupon (former Lupon District Hospital)	Dr. Raul Failma (087) 808-0122 Dr. Emmanuel Pinsoy (087) 808-0122/0920-765-0835	Townsite, Lupon, Davao Oriental									

For the purpose of this Bulletin and for better understanding of its contents, the following rules shall apply: (a) ~~Double Strike~~ – denotes deletion; (b) Underline – denotes inclusion or new item/requirement; and "xxx" – denotes separation of ~~phrases/s~~ being amended from the rest of the original text.

Philippine Road and Overseas Street, Paco, Manila, Philippines 1007

Trunk line: (02) 689 7750 | helpdesk@ps-philgeps.gov.ph



	1 unit	Compostela Valley Provincial Hospital – Montevista	Dr. Ferdinand Anthony C. Soberano (087) 808-0122 <u>Dr. Eunice B. Lawas-Manansala, MD, FPSMS</u> OIC, Head of Hospital 0917-145-7417	Montevista, Davao del Norte <u>Km. 104 Bankerohan Sur, Montevista, Compostela Valley</u>	To clarify the requirement for better understanding.
4	1 unit	Northern Mindanao Medical Center	Dr. Jose Chan (08822) 726362 <u>Dr. Jose Chan</u> (08822) 726362, 723735 local 607, 728829, (0882)721794 (TF)	Capitol Compound, Cagayan de Oro City	
5	1 unit	Northern Mindanao Medical Center	Dr. Jose Chan (08822) 726362 <u>Dr. Jose Chan</u> (08822) 726362, 723735 local 607, 728829, (0882)721794 (TF)	Capitol Compound, Cagayan de Oro City	
	1 unit	Biliran Provincial Hospital	Dr. Joyce P. Caneja (053) 500-9415	Biliran Provincial Hospital Compound <u>Biliran Provincial Hospital Compound, Naval, Biliran</u>	

SECTION VI. SCHEDULE OF REQUIREMENTS

Schedule of Requirements

page 74

LOT NO.	QTY/UOM	ITEM/DESCRIPTION	DELIVERY PERIOD
1	14 units	Blood Bank Refrigerator, 1-door, Upright, at least 50 bag capacity	Within Ninety (90) One Hundred Twenty (120)
2	8 units	Blood Bank Refrigerator, Upright, at least 350 bag capacity of 450 ml	

To adjust the delivery schedule as approved by the end-user

3	1 unit	Blood Bank Refrigerator, Upright, at least 500 bag capacity of 450 ml	Calendar Days from the effective date indicated in the Notice to Proceed
4	5 units	Blood Bank Refrigerator, 1-door, Upright, at least 700 bag capacity of 450 ml	
5	13 units	Blood Bank Refrigerator, 2-door Upright, at least 1,000 bag capacity of 450 ml	

SECTION VI. SCHEDULE OF REQUIREMENTS

SUBMISSION OF DEMO UNIT
page 75
xxx

The demo unit shall be submitted to the Technical Working Group in ~~Department of Health~~ **Procurement Service – Department of Budget and Management** for evaluation within fifteen (15) calendar days from receipt of notice from the BAC.

To amend the requirement as clarified by the end-user

SECTION VI. TECHNICAL SPECIFICATIONS

LOT NO.	AGENCY SPECIFICATION
1	Standard Features: <ul style="list-style-type: none"> • Temperature range from 2°C to 6°C Capable to set a maintained temperature range of 2°C to 6°C • Adjustable leveling feet With adjustable leveling feet for stabilization • Stainless steel/zinc plated steel wired/perforated shelves or drawers
2	
3	
4	
5	

To amend the requirement as clarified by the end-user

xxx

Lot No. 2: Blood Bank Refrigerator, Upright, at least 350 ~~350~~ 330-350 bag capacity of 450 ml
page 79
xxx

<p>Lot No. 4: Blood Bank Refrigerator, 1-door Upright, at least 700 670-700 bag capacity of 450 ml page 83 xxx</p> <p>Lot No. 5: Blood Bank Refrigerator, 2-door Upright, at least 1,000 bag capacity of 450 ml page 85 xxx</p> <p>Standard Features:</p> <ul style="list-style-type: none"> • Stainless steel drawers (10-14) shelves 	<p>To amend the requirement as clarified by the end-user</p>
--	--

B. CLARIFICATIONS

The Department of Budget and Management - Procurement Service, Bids and Awards Committee hereby clarifies prospective bidders' concerns and queries:

ITEM	CONCERN	REFERENCE	CLARIFICATION/ RESOLUTION
CONCERNS RAISED DURING PRE-BID			
1	Prospective bidders requested to allow similar contracts for the Supply and Delivery of Hospital/Medical Equipments in addition to Blood Bank Equipments	BDS Clause 5.4 page 38 xxx For this purpose, similar contracts shall refer to any: > <i>Contract for Supply and Delivery of Blood Bank Equipment</i>	Consistent with the requirement as clarified by the end-user.
2	A prospective bidder inquired if they still have to attach proof on the projects stated in the all-ongoing contracts and awarded but not yet started contracts.	BDS Clause 12.1 (a)(ii) page 41 Bidders shall submit separate statements for: (1) single largest similar completed contract/s; similar to the contract to be bid and (2) all on-going contracts and awarded but not yet started contracts.	For the Statement of All on-going contracts and awarded but not yet started contracts, just have to accomplish and submit the statement and there is no need to attach any verifiable proof. The BAC has the responsibility to validate and ascertain such statement.

3	A prospective bidder asked if there is a color coding on the bid envelope to be submitted during the deadline of submission of bids	BDS Clause 20.3 page 43 xxx Each Bidder shall submit one (1) original and one (1) copy of the first and second components of its bid	The Committee does not require specific color of envelope to be used.
4	A prospective bidder requested to reduce the warranty period of three (3) years to two (2) years for parts and service since the standard warranty given by the manufacturer is one (1) year for parts and two (2) years for service.	BDS Clause 29.2 Page 45 xxx 9. Warranty Certificate of three (3) years for parts and service, upon delivery, inspection and acceptance duly issued by the Manufacturer and Prospective Bidder.	Consistent with the requirement as clarified by the end-user.
5	A prospective bidder suggested to adjust the delivery period from ninety (90) calendar days to one hundred twenty (120) working days since their factories are loaded with work and deadlines for the 4 th quarter. And one more bidder has the same opinion on this matter.	Schedule of Requirements page 74 xxx Within Ninety (90) Calendar Days from the effective date indicated in the Notice to Proceed	Refer to the above amendment
6	A prospective bidder asked if they could just invite the Technical Working Group of the Department of Health to test their installed Blood Bank Refrigerator unit with the same catalogue number, brands, and model.	Schedule of Requirements Submission of Demo Unit page 75 xxx The demo unit shall be submitted to the Technical Working Group in Department of Health for evaluation within fifteen (15) calendar days from receipt of notice from the BAC	Refer to the above amendment
	Another bidder inquired if they can submit only one unit to represent as a sample for the five (5) lots since the item per lot only differ with their capacity		

7	<p>A prospective bidder suggested to modify the capacity of Lot 1 from 50 bags to 30-50 bags since the ABC allotted for Lot 1 is insufficient to purchase the refrigerator for at least 50 bag-capacities.</p>	<p>Technical Specification Lot 1: Blood Bank Refrigerator, 1-door, Upright, at least 50 bag capacity Page 76</p>	<p>Consistent with the requirement as clarified by the end-user.</p>
8	<p>One of the prospective bidders requested to change the temperature range to +1°C to +8°C.</p> <p>On the other hand, another bidder requested to reduce the temperature range.</p>	<p>Technical Specification <i>Applicable to all Lots</i> xxx</p> <ul style="list-style-type: none"> • Temperature range from 2°C to 6°C 	<p>Refer to the above amendment</p>
9	<p>A prospective bidder asked if the window as described in the requirement refers to the door of the refrigerator. She wanted to clarify this matter since they want to ensure that their offered item is uniformed with the given standard features of the End-User. Moreover, another bidder informed the Committee that as per World Health Organization (WHO) guidelines, there is no specification that it has to be in an adjustable levelling feet. The guidelines only requires that the equipment must be movable and the wheels must be able to be locked.</p> <p>Furthermore, one of the prospective bidders suggested that as an alternative to stainless steel wire shelves they can offer stainless steel wire drawers.</p> <p>A supplier, likewise, inquired if it is possible that the accessible interior light switch may be located outside the control panel.</p>	<p>Technical Specification <i>Applicable to all Lots</i> xxx</p> <ul style="list-style-type: none"> • Insulated steel frame w/ two (2) layer glass windows w/ lock • Stainless steel/zinc plated steel wireshelves <p>xxx</p> <ul style="list-style-type: none"> • Accessible interior light switch • Adjustable leveling feet 	<p>The End-User explained that in fact the word "accessible" in the requirement means that it must be located outside the control panel but lights the interior of the equipment.</p> <p>As regards with the description of the two (2) layer glass windows, the End-User explained that the door of the refrigerator must be covered with two (2) layers of glass. Double pane glass windows can have a gas fill between the panes of glass for added thermal efficiency.</p>

10	A representative from EMI and SERO Diagnostic Trading requested to modify the description of 1-door to a 2-door blood bank refrigerator since the capacity required in Lot No. 4 is at least 700 bags.	Technical Specification Lot No. 4: Blood Bank Refrigerator , 1-door Upright, at least 700 bag capacity of 450 ml page 83	Refer to the above amendment
11	A prospective bidder suggested to lessen the bag capacity from 1000 to a range of 700-1000.	Technical Specification LOT NO. 5 : Blood Bank Refrigerator , 2-door Upright, at least 1,000 bag capacity of 450 ml Page 85	Consistent with the requirement as clarified by the end-user.
12	A prospective bidder, recommended the adjustment in the number of shelves for the steel drawers. The suggestion was to have a range of 12-14 shelves.	Technical Specification LOT NO. 5 : Blood Bank Refrigerator , 2-door Upright, at least 1,000 bag capacity of 450 ml Page 85 xxx Standard Features • Stainless steel drawers (14) shelves	Refer to the above amendment.
13	A prospective bidder said that as a substitute for a Built-in Circuit Breaker, an external circuit breaker be considered	Technical Specification LOT NO. 5 : Blood Bank Refrigerator , 2-door Upright, at least 1,000 bag capacity of 450 ml Page 86 xxx Safety Features: Audible/Visual Alarms: • Built in Circuit Breaker	Consistent with the requirement as clarified by the end-user.

ITEM	CONCERN	REFERENCE	CLARIFICATION/ RESOLUTION
LABTRADERS, INC.			
1	May we request for the temperature range to be adjusted to 1-8C from 2-6C. The range 1-8C actually covers the required temperature range	Technical Specification: <ul style="list-style-type: none"> • Temperature range from 2°C to 6°C 	Refer to the above amendment.
2	For Lot 3 - May we request to adjust the required blood bag capacity from 500 bags to a <u>minimum of 425 bags</u> ;	Technical Specification Page 80 xxx At least 500 bag capacity of 450 mL	Consistent with the requirement as clarified by the end-user.
3	For Lot 4 - May we request to accept as equivalent to the requirement of two (2) layer glass window with lock: <u>two (2) window pane glass door with lock</u> ;	Technical Specification Page 82 <ul style="list-style-type: none"> • Insulated steel frame w/ two (2) layer glass windows w/ lock 	Consistent with the requirement as clarified by the end-user.
4	For Lot 5 – May we request to adjust the required blood bag capacity from 1000 blood bags to a <u>minimum of 750 blood bags</u> ; May we request to also accept <u>External Circuit Breaker</u> ;	Technical Specification Page 84 xxx <ul style="list-style-type: none"> • At least 1000 bag capacity of 450 ml • Built in Circuit Breaker 	Consistent with the requirement as clarified by the end-user.
5	May we request to accept as equivalent to the requirement of adjustable levelling feet: four (4) caster wheels, front two (2) casters lockable for easy mobility	Technical Specification: For all the Lots. xxx <ul style="list-style-type: none"> • Adjustable leveling feet 	Refer to the above amendment.
6	With regards to the <u>requirement for demo units</u> , may we request that on-site installation visit be accepted as an option	Submission of Demo Unit page 74 (a) xxx The demo unit shall be submitted to the TWG in Department of Health within (15) calendar days	Refer to the above amendment.

7	With regards to Single Largest Contract for Lot 5, may we request for the BAC to accept two (2) similar contracts with an aggregate amount equivalent to, at least, 50% of the ABC of the Lot	Bid Data Sheet Clause 5.4 page 40 xxx The bidder must have completed within the period specified in the Invitation to bid, a single contract that is similar to this Project, equivalent to at least 50% of the ABC	Consistent with the requirement as clarified by the end-user.
ENDURE MEDICAL, INC.			
1	On Lot No. 2 – will the BAC consider changing the requirement from "Stainless steel/zinc plated steel wire shelves" into "Stainless steel/zinc plated steel wire shelves or stainless-steel drawers " in order to encourage more bidders to participate?	Technical Specification: Lot 2: At least 350 bag capacity page 78 xxx > Stainless steel/zinc plated steel wired/perforated shelves	Refer to the above amendment.
2	On Lot No. 4 - will the BAC remove the requirement "1-door" in order to allow participation of companies which offer "two-doors" Blood Bank Refrigerators? Will the BAC also consider changing the requirement from "Stainless steel/zinc plated steel wire shelves" into "Stainless steel/zinc plated steel wire shelves or stainless steel drawers '?	Technical Specification: Lot 4: At least 700 bag capacity page 82 xxx Blood Bank Refrigerator, 1-door , Upright, at least 700 bag capacity of 450 ml Stainless steel/zinc plated steel wired/perforated shelves	Refer to the above amendment.
3	On Lot No. 5 - will the BAC accept an offer of equipment that has only "at least 720 bag capacity" instead of 1000 bag capacity? Will the BAC also consider changing the requirement from Stainless steel drawers (14) shelves into from "Stainless steel drawers (12) shelves" to allow more bidders to participate?	Technical Specification: Lot 5: At least 1,000 bag capacity page 84 xxx Blood Bank Refrigerator, 2-door, Upright, at least 1,000 bag capacity of 450 ml Stainless steel drawers (14) shelves	Consistent with the requirement as clarified by the end-user Stainless steel drawers (10 to 14) shelves
MEDICAL CENTER TRADING CORPORATION			

1	<p>Requesting to allow similar contracts of Supply and Delivery of <u>Hospital/Medical and Laboratory Equipments</u> in addition to Blood Bank Equipments.</p>	<p>Bid Data Sheet Clause 5.4 page 40 xxx For this purpose, similar contracts shall refer to any:</p> <ul style="list-style-type: none"> • <i>Contract for Supply and Delivery of Blood Bank Equipment</i> 	<p>Consistent with the requirement as clarified by the end-user</p>
---	---	---	---

All other portions of the Bidding Documents affected by these amendments shall be made to conform to the same.

Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

(SGD)ARIEL R. CUNANAN
Chairperson, PS BAC V

Schedule of Requirements

LOT NO.	QTY/UOM	ITEM/DESCRIPTION	DELIVERY PERIOD
1	14 units	Blood Bank Refrigerator , 1-door, Upright, at least 50 bag capacity	Within <u>One Hundred Twenty (120)</u> Calendar Days from the effective date indicated in the Notice to Proceed
2	8 units	Blood Bank Refrigerator , Upright, at least 350 bag capacity of 450 ml	
3	1 unit	Blood Bank Refrigerator , Upright, at least 500 bag capacity of 450 ml	
4	5 units	Blood Bank Refrigerator , 1-door, Upright, at least 700 bag capacity of 450 ml	
5	13 units	Blood Bank Refrigerator , 2-door Upright, at least 1,000 bag capacity of 450 ml	

I hereby commit to comply and delivery all the above requirements in accordance with the above stated schedule.

Name of Company	Signature Over Printed Name of Authorized Representative	Date
-----------------	---	------

TECHNICAL SPECIFICATIONS

LOT NO. 1 : **Blood Bank Refrigerator**, 1-door, Upright, at least
 : 50 bag capacity
QUANTITY : **14 units**
APPROVED BUDGET FOR THE CONTRACT : **P 3,500,000.00**

AGENCY SPECIFICATIONS	BIDDER'S STATEMENT OF COMPLIANCE	REFERENCE
Power Supply Input: <ul style="list-style-type: none"> • 220-240 VAC, 60Hz 	Brand and Model:	
Standard Features: <ul style="list-style-type: none"> • <u>Capable to set a maintained temperature range of 2°C to 6°C</u> • CFC – free insulation • CFC – free refrigerant • Rigid polyurethane cabinet insulation • Insulated steel frame w/ two (2) layer glass windows w/ lock • Stainless steel/zinc plated steel wired/perforated shelves <u>or drawers</u> • Temperature Display • Circulating Fan • Temperature control • Light Door activated • Accessible interior light switch • <u>With adjustable leveling feet for stabilization</u> 		
Safety Features: Audible/Visual Alarms: <ul style="list-style-type: none"> • For High & Low temperature • Door ajar alarm • Power cut-off 		

<ul style="list-style-type: none"> • Push button alarm tests • Alarm silence switch 		
<p>Standard Accessories:</p> <ul style="list-style-type: none"> • Automatic Voltage regulator (Compatible to the equipment power requirement) 		
<p>Packaging Instruction:</p> <ul style="list-style-type: none"> • Standard packaging of the manufacturer 		
<p>Labeling Instruction:</p> <ul style="list-style-type: none"> • Each Blood Bank Refrigerator, 1-door Upright, at least 50 bag capacity must have non-removable / permanent markings on visible area "Philippine Government Property – Department of Health – Not for Sale", 1-inch in size on front surface. 		

I hereby certify that the statement of compliance to the foregoing technical specifications are true and correct, otherwise, if found to be false either during bid evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.

Name of Company	Signature Over Printed Name of Authorized Representative	Date
-----------------	---	------

LOT NO. 2

- **Blood Bank Refrigerator**, Upright, 330-350 bag capacity of 450 ml

QUANTITY

: **8 units**

APPROVED BUDGET FOR THE CONTRACT

: **P 5,600,000.00**

AGENCY SPECIFICATIONS	BIDDER'S STATEMENT OF COMPLIANCE	REFERENCE
Power Supply Input: <ul style="list-style-type: none">• 220-240 VAC, 60Hz	Brand and Model:	
Standard Features: <ul style="list-style-type: none">• <u>Capable to set a maintained temperature range of 2°C to 6°C</u>• CFC – free insulation• CFC – free refrigerant• Rigid polyurethane cabinet insulation• Insulated steel frame w/ two (2) layer glass windows w/ lock• Stainless steel/zinc plated steel wired/perforated shelves <u>or drawers</u>• Temperature Display• Circulating Fan• Temperature control• Light Door activated• Accessible interior light switch• <u>With adjustable leveling feet for stabilization</u>		
Safety Features: Audible/Visual Alarms: <ul style="list-style-type: none">• For High & Low temperature• Door ajar alarm• Power cut-off• Alarm test Capable• Alarm silence switch		

<p>Standard Accessories:</p> <ul style="list-style-type: none"> • Automatic Voltage regulator (Compatible to the equipment power requirement) 		
<p>Packaging Instruction:</p> <ul style="list-style-type: none"> • Standard packaging of the manufacturer 		
<p>Labeling Instruction:</p> <ul style="list-style-type: none"> • Each Blood Bank Refrigerator, 1-door, at least 350 bag capacity of 450 ml must have non-removable / permanent markings on visible area "Philippine Government Property – Department of Health – Not for Sale", 1-inch in size on front surface. 		

I hereby certify that the statement of compliance to the foregoing technical specifications are true and correct, otherwise, if found to be false either during bid evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.

Name of Company

Signature Over Printed Name
of Authorized Representative

Date

LOT NO. 3

- **Blood Bank Refrigerator**, Upright, at least 500 bag capacity of 450 ml

QUANTITY

: 1 unit

APPROVED BUDGET FOR THE CONTRACT

: P 650,000.00

AGENCY SPECIFICATIONS	BIDDER'S STATEMENT OF COMPLIANCE	REFERENCE
Power Supply Input: <ul style="list-style-type: none">• 220-240 VAC, 60Hz	Brand and Model:	
Standard Features: <ul style="list-style-type: none">• <u>Capable to set a maintained temperature range of 2°C to 6°C</u>• CFC – free insulation• CFC – free refrigerant• Rigid polyurethane cabinet insulation• Insulated steel frame w/ two (2) layer glass windows w/ lock• Stainless steel/zinc plated steel wired/perforated shelves <u>or drawers</u>• Temperature Display• Circulating Fan• Temperature control• Light Door activated• Accessible interior light switch• <u>With adjustable leveling feet for stabilization</u>		
Safety Features: Audible/Visual Alarms: <ul style="list-style-type: none">• For High & Low temperature• Door ajar alarm• Power cut-off• Alarm test Capable• Alarm silence switch		

<p>Standard Accessories:</p> <ul style="list-style-type: none"> • Automatic Voltage regulator (Compatible to the equipment power requirement) 		
<p>Packaging Instruction:</p> <ul style="list-style-type: none"> • Standard packaging of the manufacturer 		
<p>Labeling Instruction:</p> <ul style="list-style-type: none"> • Each Blood Bank Refrigerator, Upright, at least 500 bag capacity of 450 ml must have non-removable / permanent markings on visible area "Philippine Government Property – Department of Health – Not for Sale", 1-inch in size on front surface. 		

I hereby certify that the statement of compliance to the foregoing technical specifications are true and correct, otherwise, if found to be false either during bid evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.

Name of Company

Signature Over Printed Name
of Authorized Representative

Date

LOT NO. 4

- **Blood Bank Refrigerator**, Upright, 670-700 bag capacity of 450 ml

QUANTITY

: 5 units

APPROVED BUDGET FOR THE CONTRACT

: P 3,500,000.00

AGENCY SPECIFICATIONS	BIDDER'S STATEMENT OF COMPLIANCE	REFERENCE
Power Supply Input: <ul style="list-style-type: none">• 220-240 VAC, 60Hz	Brand and Model:	
Standard Features: <ul style="list-style-type: none">• <u>Capable to set a maintained temperature range of 2°C to 6°C</u>• CFC – free insulation• CFC – free refrigerant• Rigid polyurethane cabinet insulation• Insulated steel frame w/ two (2) layer glass windows w/ lock• Stainless steel/zinc plated steel wired/perforated shelves <u>or drawers</u>• Temperature Display• Circulating Fan• Temperature control• Light Door activated• Accessible interior light switch• <u>With adjustable leveling feet for stabilization</u>		
Safety Features: Audible/Visual Alarms: <ul style="list-style-type: none">• For High & Low temperature• Door ajar alarm• Power cut-off• Alarm test Capable• Alarm silence switch		

<p>Standard Accessories:</p> <ul style="list-style-type: none"> • Automatic Voltage regulator (Compatible to the equipment power requirement) 		
<p>Packaging Instruction:</p> <ul style="list-style-type: none"> • Standard packaging of the manufacturer 		
<p>Labeling Instruction:</p> <ul style="list-style-type: none"> • Each Blood Bank Refrigerator, 1-door, at least 700 bag capacity of 450 ml must have non-removable / permanent markings on visible area "Philippine Government Property – Department of Health – Not for Sale", 1-inch in size on front surface. 		

I hereby certify that the statement of compliance to the foregoing technical specifications are true and correct, otherwise, if found to be false either during bid evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.

Name of Company	Signature Over Printed Name of Authorized Representative	Date
-----------------	---	------

LOT NO. 5

- **Blood Bank Refrigerator**, 2-door Upright, at least 1,000
- bag capacity of 450 ml

QUANTITY

: 13 units

APPROVED BUDGET FOR THE
CONTRACT

: P 17,550,000.00

AGENCY SPECIFICATIONS	BIDDER'S STATEMENT OF COMPLIANCE	REFERENCE
Power Supply Input: <ul style="list-style-type: none">• 220-240 VAC, 60Hz	Brand and Model:	
Standard Features: <ul style="list-style-type: none">• Controller Interface• Circulating Fan• Communication ports for data retrieval• With digital temperature display• <u>Capable to set a maintained temperature range of 2°C to 6°C</u>• With temperature recording device• CFC – free insulation• CFC – free refrigerant• Rigid polyurethane cabinet insulation• Double Door, Insulated steel frame• Door Construction: Glass dual-pane glass doors• Stainless steel drawers (<u>10-14</u>) shelves• Accessible interior/exterior light switch• Accuracy of settings: Chamber temperature is +/- 1°C (Temperature difference between upper and lower chamber)		
Safety Features: Audible/Visual Alarms: <ul style="list-style-type: none">• Built in Circuit Breaker		

<ul style="list-style-type: none"> • For High & Low temperature • Door ajar alarm • Power cut-off • Push button alarm tests • Alarm silence switch • Rechargeable Battery Back-up system for at least 20 hours. 		
<p>Standard Accessories:</p> <ul style="list-style-type: none"> • Automatic Voltage regulator (Compatible to the equipment power requirement) 		
<p>Packaging Instruction:</p> <ul style="list-style-type: none"> • Standard packaging of the manufacturer 		
<p>Labeling Instruction:</p> <ul style="list-style-type: none"> • Each Blood Bank Refrigerator, 2-door, at least 1,000 bag capacity of 450 ml must have non-removable / permanent markings on visible area "Philippine Government Property – Department of Health – Not for Sale", 1-inch in size on front surface. 		

I hereby certify that the statement of compliance to the foregoing technical specifications are true and correct, otherwise, if found to be false either during bid evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.

Name of Company	Signature Over Printed Name of Authorized Representative	Date
-----------------	---	------