

Republic of the Philippines Department of Budget and Management PROCUREMENT SERVICE BIDS AND AWARDS COMMITTEE



Bid Bulletin No.1 12 March 2019

PUBLIC BIDDING NO. 19-075-10

SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF COMPUTED TOMOGRAPHY (CT) SCANNER (16 SLICE) FOR ZAMBOANGA CITY MEDICAL CENTER (ZCMC)

Issued pursuant to Sec. 22.5 of the IRR of R.A. 9184 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by prospective bidder/s during the Pre-Bid Conference held on **March 11, 2019**.

A. AMENDMENTS/INCLUSIONS

REFERENCE		BASIS FOR AMENDMENT/INCLUSIONS
SECTION I. INVITATION	I TO BID	
Page 4		
xxx		
2. The summary of the bidd	ling activities is as follows:	
Advertisement/Posting of Invitation to Bid	March 04, 2019	To give the prospective bidders ample
Issuance and Availability of Bid Documents	March 04, 2019	time to prepare for their clarifications.
Pre-Bid Conference	March 11, 2019; 10:00 AM	
Last day of Submission of Written Clarification	March 13 <u>14</u> , 2019	
Last day of Issuance of Bid Bulletin	March 18, 2019	
Deadline for Submission	March 25, 2019; 10:00 AM	

xxx		
SECTION III. BID DATA SHEET		
Page 43		
xxx		
3. Actual product presentation of the same brand and model of the CT Scanner being offered. Commitment letter that the bidder shall facilitate a site visit of their actual installation of CT Scanner of the same brand and model being offered including product presentation upon the request of BAC X.	To make the requirement specific for better understanding of prospective bidders.	
xxx		
SECTION V. SPECIAL CONDITIONS OF THE CONT	RACT	
SCC Clause 11.3		
Page 66		
Terms of Payment:		
10% - Mobilization – upon issuance of the contract 65% - Upon Delivery, Installation, Testing and Commissioning. 25% - Upon acceptance of the Joint PS-MRH's ZCMC's Inspection Team and approval of the CDRRHR-FDA.	Typographical Error	
SECTION VII. TECHNICAL SPECIFICATIONS		
Page 74		
THIRD DARTY ACCESSORIES		
THIRD PARTY ACCESSORIES	• To give options to prospective	
XXX		
Contrast Injector- single/dual barrel compatible with the CT Scan unit with 500 pieces of syringe sets compatible with the Single/Dual Barrel Power Injector		

Page 74

XXX

ADDITIONAL REQUIREMENTS:

1. Three (3) years warranty on parts (including X-ray tube) and services for the CT Scanner and all <u>its</u> accessories <u>including the third part accessories</u>. Warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.

XXX

3. Supplier must specify post warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the CT Scanner and all <u>its</u> accessories <u>including the third part accessories</u> for the next three (3) years after the warranty period.

XXX

 To clearly state that the third-party accessories are included under the coverage of the required warranty.

By Rule, all other related provisions in the Bidding Documents correspondingly affected by these amendments are likewise deemed amended to conform to this Bid Bulletin.

Amendments made herein shall be considered an integral part of the Bidding Documents.

(SGD) ENGR. PABLO ROMAN C. ANDRESChairperson, Bids and Awards Committee X

For the purpose of this Bulletin and for better understanding of its contents, the following rules shall apply: (a) double Strikeout – denotes deletion; (b) <u>Underline</u> – denotes inclusion or new item/requirement; and "xxx" – denotes separation of phase/s being amended from the rest of the main text.