



Republic of the Philippines
 Department of Budget and Management
PROCUREMENT SERVICE
BIDS AND AWARDS COMMITTEE



Bid Bulletin No.1
12 April 2019

PUBLIC BIDDING NO. 19-114-10

**PROCUREMENT OF MEDICAL AND LABORATORY EQUIPMENT FOR THE
 LAS PIÑAS GENERAL HOSPITAL SATELLITE TRAUMA CENTER (LPGH-STC)**

Issued pursuant to Sec. 22.5 of the IRR of R.A. 9184 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by prospective bidder/s during the Pre-Bid Conference held on **10 April 2019**.

A. AMENDMENTS/INCLUSIONS

REFERENCE	BASIS FOR AMENDMENT/INCLUSIONS										
Section III BID DATA SHEET											
<p>BDS Clause 18.1: Page 40</p> <p>xxx</p> <p>The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts:</p> <table border="1" data-bbox="135 1491 997 1825"> <thead> <tr> <th>Lot No.</th> <th>Item/Description</th> <th>Amount Cash, Cashier's / Manager's Check, Bank Draft/ Guarantee/Irrevocable Letter of Credit (2%)</th> <th>Surety Bond (5%)</th> <th>Bid Securing Declaration (Pursuant to GPPB Resolution No. 03- 2012)</th> </tr> </thead> <tbody> <tr> <td>21</td> <td>SEROFUGE/SEROLOGICAL CENTRIFUGE</td> <td>5,769.23</td> <td>14,423.08</td> <td>Please see</td> </tr> </tbody> </table>	Lot No.	Item/Description	Amount Cash, Cashier's / Manager's Check, Bank Draft/ Guarantee/Irrevocable Letter of Credit (2%)	Surety Bond (5%)	Bid Securing Declaration (Pursuant to GPPB Resolution No. 03- 2012)	21	SEROFUGE/SEROLOGICAL CENTRIFUGE	5,769.23	14,423.08	Please see	<p>➤ Typographical Error</p>
Lot No.	Item/Description	Amount Cash, Cashier's / Manager's Check, Bank Draft/ Guarantee/Irrevocable Letter of Credit (2%)	Surety Bond (5%)	Bid Securing Declaration (Pursuant to GPPB Resolution No. 03- 2012)							
21	SEROFUGE/SEROLOGICAL CENTRIFUGE	5,769.23	14,423.08	Please see							



2	DYSPHAGIA THERAPY WITH EMG BIOFEEDBACK	6,730.77	16,826.92	attached "Annex H"
4-3	FRACTURE / ORTHO SURGICAL TABLE (AUTOMATIC/MOTORIZED WITH TRACTION ATTACHMENT)	28,846.15	72,115.38	

Section VI. TECHNICAL SPECIFICATIONS

Page 72

XXX

Lot No.	Items and Description
1	<p>SEROFUGE/SEROLOGICAL CENTRIFUGE</p> <p>Specifications:</p> <p>XXX</p> <p>-Maximum speed RPM: 10,000 – 15,000 rpm or greater</p> <p>XXX</p>
2	<p>DYSPHAGIA THERAPY WITH EMG BIOFEEDBACK</p> <p>Specifications:</p> <p>XXX</p> <p>OLED Screen LED / OLED Screen</p> <p>XXX</p> <p>XXX</p> <p>Low battery warning function</p> <p>XXX</p> <p>XXX</p> <p>Rechargeable lithium battery & micro-USB port Connected to Microsoft PC by USB cable (Power</p>

➤ Modifying Technical Specifications of the End-user to give more opportunities to other bidders

	supplied directly by PC)	
	XXX	

B. CLARIFICATIONS

DETAIL	CLARIFICATION				
Letter from Jordal Medical Systems, Inc.¹					
Section VII Technical Specifications					
Page 73					
<table border="1" style="width: 100%;"> <thead> <tr> <th style="width: 10%;">Lot No.</th> <th style="width: 90%;">Items and Description</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">2</td> <td> DYSPHAGIA THERAPY WITH EMG BIOFEEDBACK Specifications: XXX OLED Screen XXX XXX -Low battery warning function Xxx XXX - Rechargeable lithium battery & micro USB port XXX </td> </tr> </tbody> </table>	Lot No.	Items and Description	2	DYSPHAGIA THERAPY WITH EMG BIOFEEDBACK Specifications: XXX OLED Screen XXX XXX -Low battery warning function Xxx XXX - Rechargeable lithium battery & micro USB port XXX	<ul style="list-style-type: none"> ➤ Suggest to change it to LED Screen ➤ Delete "Low battery warning function" ➤ Suggest to modify it to "connected to Microsoft PC by USB cable (Power supplied directly by PC ➤ In reference to the letter of LPGH-STC² dated 11 April 2019, the End-User are amenable to the said changes
Lot No.	Items and Description				
2	DYSPHAGIA THERAPY WITH EMG BIOFEEDBACK Specifications: XXX OLED Screen XXX XXX -Low battery warning function Xxx XXX - Rechargeable lithium battery & micro USB port XXX				

By Rule, all other related provisions in the Bidding Documents correspondingly affected by these amendments are likewise deemed amended to conform to this Bid Bulletin.

¹ Email letter dated 10 April 2019

² Email letter dated 15 April 2019

Amendments made herein shall be considered an integral part of the Bidding Documents.

(SGD) ENGR. PABLO ROMAN C. ANDRES
Chairperson, PS BAC X

For the purpose of this Bulletin and for better understanding of its contents, the following rules shall apply: (a) ~~Double Strike out~~ — denotes deletion; (b) Underline — denotes inclusion or new item/requirement; and "xxx" — denotes separation of phrase/s being amended from the rest of the main text.

	Attached as Annexes "C" and "C-1" in Section VIII. Bidding Forms are the standard forms for this requirement including the instructions and guidelines in the accomplishment of said forms.																			
13.1(b)	No further instructions.																			
13.1(c)	No additional requirements.																			
13.2	The ABC are as follows:																			
	<table border="1"> <thead> <tr> <th>Lot No.</th> <th>Quantity</th> <th>Item/Description</th> <th>Approved Budget for the Contract</th> </tr> </thead> <tbody> <tr> <td>1</td> <td>1 Unit</td> <td>SEROFUGE / SEROLOGICAL CENTRIFUGE</td> <td>288,461.54</td> </tr> <tr> <td>2</td> <td>1 Unit</td> <td>DYSPHAGIA THERAPY WITH EMG BIOFEEDBACK</td> <td>336,538.46</td> </tr> <tr> <td>3</td> <td>1 Unit</td> <td>FRACTURE / ORTHO SURGICAL TABLE (AUTOMATIC/MOTORIZED WITH TRACTION ATTACHMENT)</td> <td>1,442,307.69</td> </tr> </tbody> </table>				Lot No.	Quantity	Item/Description	Approved Budget for the Contract	1	1 Unit	SEROFUGE / SEROLOGICAL CENTRIFUGE	288,461.54	2	1 Unit	DYSPHAGIA THERAPY WITH EMG BIOFEEDBACK	336,538.46	3	1 Unit	FRACTURE / ORTHO SURGICAL TABLE (AUTOMATIC/MOTORIZED WITH TRACTION ATTACHMENT)	1,442,307.69
Lot No.	Quantity	Item/Description	Approved Budget for the Contract																	
1	1 Unit	SEROFUGE / SEROLOGICAL CENTRIFUGE	288,461.54																	
2	1 Unit	DYSPHAGIA THERAPY WITH EMG BIOFEEDBACK	336,538.46																	
3	1 Unit	FRACTURE / ORTHO SURGICAL TABLE (AUTOMATIC/MOTORIZED WITH TRACTION ATTACHMENT)	1,442,307.69																	
	Any bid with a financial component exceeding this amount shall not be accepted.																			
15.4(a)(iv)	No incidental services are required.																			
15.4(b)	No incidental services are required.																			
16.1(b)	The Bid prices for Goods supplied from outside of the Philippines shall be quoted in Philippine Pesos.																			
16.3	Not applicable.																			
17.1	Bids will be valid for one hundred twenty (120) calendar days from opening of bids.																			
18.1	The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts:																			
	<table border="1"> <thead> <tr> <th>Lot No.</th> <th>Item/Description</th> <th>Amount Cash, Cashier's / Manager's Check, Bank Draft/ Guarantee/Irrevo- cable Letter of Credit (2%)</th> <th>Surety Bond (5%)</th> <th>Bid Securing Declaration (Pursuant to GPPB Resolution No. 03-2012)</th> </tr> </thead> <tbody> <tr> <td>1</td> <td>SEROFUGE/SEROLOGICAL CENTRIFUGE</td> <td>5,769.23</td> <td>14,423.08</td> <td rowspan="2">Please see attached "Annex H"</td> </tr> <tr> <td>2</td> <td>DYSPHAGIA THERAPY WITH EMG BIOFEEDBACK</td> <td>6,730.77</td> <td>16,826.92</td> </tr> </tbody> </table>				Lot No.	Item/Description	Amount Cash, Cashier's / Manager's Check, Bank Draft/ Guarantee/Irrevo- cable Letter of Credit (2%)	Surety Bond (5%)	Bid Securing Declaration (Pursuant to GPPB Resolution No. 03-2012)	1	SEROFUGE/SEROLOGICAL CENTRIFUGE	5,769.23	14,423.08	Please see attached "Annex H"	2	DYSPHAGIA THERAPY WITH EMG BIOFEEDBACK	6,730.77	16,826.92		
Lot No.	Item/Description	Amount Cash, Cashier's / Manager's Check, Bank Draft/ Guarantee/Irrevo- cable Letter of Credit (2%)	Surety Bond (5%)	Bid Securing Declaration (Pursuant to GPPB Resolution No. 03-2012)																
1	SEROFUGE/SEROLOGICAL CENTRIFUGE	5,769.23	14,423.08	Please see attached "Annex H"																
2	DYSPHAGIA THERAPY WITH EMG BIOFEEDBACK	6,730.77	16,826.92																	

	43	FRACTURE / ORTHO SURGICAL TABLE (AUTOMATIC/MOTORIZED WITH TRACTION ATTACHMENT)	28,846.15	72,115.38	
		The bid security in the form of cashier's/manager's check shall be payable to PROCUREMENT SERVICE .			
18.2		The bid security shall be valid for one hundred twenty (120) calendar days from opening of bids.			
20.3		Each Bidder shall submit one (1) original and one (1) copy of the first and second components of its bid. The duplicate- <i>i.e. copy 1</i> , must include the same documents as that of the original set of documents. In case, however, a bidder opts to submit cash as bid security, copy 1 need not contain photocopies of the same.			
21		The address for submission of bids is: <i>Bids and Awards Committee X Procurement Service Secretariat Office 2nd Floor, PS Complex, RR Road Cristobal St., Paco, Manila</i> The deadline for submission of bids is on 24 April 2019, 10:00AM			
24.1		The place of bid opening is: <i>PS Conference Room New PS Building, PS-DBM Compound, RR Road Cristobal St., Paco, Manila</i> The bid opening will be held immediately after the deadline of submission of bids. During the opening of bids only the authorized representative shall be allowed to assist in the opening of bids. In case the authorized representative is not present, any representative of the authorized representative may be allowed to assist in the opening of bids provided that a separate written authorization from the authorized representative shall be presented for the purpose. Provided that the authorized representative is duly authorized to issue such further authority and the same is reflected in the Board Resolution and/or Secretary's Certificate.			
24.2		No further instructions.			
24.3		No further instructions.			
27.1		No further instructions.			

Section VII. Specifications

LOT NO.1 : SEROFUGE/SEROLOGICAL CENTRIFUGE

QUANTITY : 1 UNIT

APPROVED BUDGET FOR THE CONTRACT : ₱ 288,461.54

Lot No.	Items and Description	UOM	Bidder's Statement of Compliance	Reference
3	<p>SEROFUGE/SEROLOGICAL CENTRIFUGE</p> <p>Specifications: -No. of plaser: 12 plaser (12 x 75 mm tube) -Maximum speed RPM: 10,000 – 15,000 rpm or greater -Maximum RCF x g: 20,376 – 25,400 g -Rotor: Fixed Angle -Motor Type: Induction or Brushless -Control type: Digital/Microcomputer -Running Time: 1s to 99 minutes -Power supply: 220V, 60 Hz</p> <p>Warranty: One (1) year on parts and Two (2) years on services</p>	1 UNIT	BRAND and MODEL:	

I hereby certify that the statement of compliance to the foregoing technical specifications are true and correct, otherwise if found false either during evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.

Name of Company

Signature over Printed Name
of Authorized Representative

Date

LOT NO. 2

DYSPHAGIA THERAPY WITH
EMG BIOFEEDBACK

QUANTITY

1 UNIT

APPROVED BUDGET FOR THE CONTRACT

₱ 336,538.46

Lot No.	Items and Description	UOM	Bidder's Statement of Compliance	Reference
2	<p>DYSPHAGIA THERAPY WITH EMG BIOFEEDBACK</p> <p>Specifications:</p> <ul style="list-style-type: none"> -LED / OLED Screen -ETS (EMG Triggered Stimulation Function) -NMES (Neuromuscular Electrical Stimulation) -Functional electrical stimulation prescription management -Electrode Isolation Technology: ETS and NMES mode use 1 Electrode Wire - Electrodes "Loose Detection" Function - Low battery warning function - Symmetric Biphasic balanced wave - Rechargeable Lithium Battery & micro USB port <u>Connected to Microsoft PC by USB cable; power supplied directly to PC</u> - Data storage function - Communication with Computer Windows 10 Laptop with standard or good graphic acceleration Suggested system: At least Intel core i3 2GB RAM minimum 500GB HDD, with USB ports <p>Warranty: One (1) year on parts and Two (2) years on services</p>	1 UNIT	BRAND and MODEL:	

I hereby certify that the statement of compliance to the foregoing technical specifications are true and correct, otherwise if found false either during evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.

Name of Company

Signature over Printed Name
of Authorized Representative

Date