



Bid Bulletin No. 1

March 05, 2020

PUBLIC BIDDING NO. 20-013-5

**SUPPLY AND DELIVERY OF MEDICAL AND LABORATORY EQUIPMENT
 FOR THE BUREAU OF QUARANTINE**

Issued pursuant to Sec. 22.5 of the IRR of R.A. 9184 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by prospective bidders during the **Pre-Bid Conference** held on **27 February 2020**, likewise, respond to bidders' written queries received within the prescriptive period for filing.

A. AMENDMENTS

REFERENCE						BASES FOR AMENDMENT
Invitation to Bid Page 4 xxx						<i>To clarify the requirement.</i>
Lot No.	QTY/ UOM	Item/ Description	Approved Budget per unit for the Contract	Price of Bid Documents	Delivery Period	
xxx						
Section III. Bid Data Sheet BDS Clause 29.2 Page 45 xxx 11. Supplier's Certificate of undertaking to conduct all the required technical training on the <u>BOQ Office at Port Area, Manila</u> as specified in Section VII. Technical Specification.						<i>To amend the training as clarified by End-User</i>
xxx						
Section VI. Schedule of Requirements Page 72 xxx						<i>To amend the Delivery Period as clarified by End-User</i>
LOT NO.	QTY/ UOM	ITEM/ DESCRIPTION	DELIVERY PERIOD			

1	2 units	Analytical Balance	<p>Within Seventy-Five (75) SIXTY (60) Calendar Days from the effective date indicated in the Notice to Proceed</p>
2	1 unit	Animal Containment Workstation	
3	2 units	Colony Counter, Automated	
4	14 units	External Defibrillator, Automated	
5	2 units	Biological Refrigerator	
6	2 units	Hot Plate Stirrer	
7	3 units	Weighing Scale, Digital	
8	4 units	Loop Sterilizer	
9	4 units	pH Meter, Portable	
10	4 units	Refrigerator, Household 2 door	
11	2 units	Stomacher	
12	2 units	Trinocular Microscope with Monitor	

Section VI. Schedule of Requirements

Page 73

xxx

Submission of Demo Unit

Bidder shall submitted the demo units to the Joint TWG of DOH BOQ-PS DBM at the PS-DBM Complex for the hereunder lots at the supplier's warehouse within fifteen (15) calendar days after receipt of notice of Single/ Lowest Calculated Bid from the BAC. Alternatively, conduct of a demo in a site within Metro Manila of an existing unit with the same brand and model delivered by the supplier is acceptable.

xxx

To clarify the requirement for Submission of Demo Unit

<p>Should the Demo unit pass the post qualification, the same shall form part of the delivery. <u>Provided it is certified by the Joint Technical Working Group to be brand new during post-qualification. Provided Likewise</u>, all consumables and parts used during test shall be replaced.</p> <p>xxx</p>	<p><i>To clarify the requirement</i></p>
<p>Section VII. Technical Specifications Page 77 Lot 2: Animal Containment Workstation</p> <p>xxx</p> <p><u>Dimensions :</u></p> <ul style="list-style-type: none"> ❖ <u>Width: 1219-1524mm</u> ❖ <u>Depth: 800-820mm</u> ❖ <u>Height: 1371-1676mm</u> <p>xxx</p>	<p><i>To include the dimension as clarified by End-User</i></p>
<p>Section VII. Technical Specifications Page 78 Lot 3: COLONY COUNTER, AUTOMATED</p> <p>xxx</p> <ul style="list-style-type: none"> • <u>Minimum size of detected colony: 0.1mm</u> <p>xxx</p> <ul style="list-style-type: none"> • <u>Compatible for all Petri dishes: 55, 65, 90, 150mm, petrifilm, and Compact Dry.</u> <p>xxx</p> <p>Accessories:</p> <ul style="list-style-type: none"> • Marking Probe • Magnifying Lens • Lens Holder <p>xxx</p>	<p><i>To include in the Technical Specifications as clarified by End-User</i></p> <p><i>To Delete the requirement as clarified by End-User</i></p>
<p>Section VII. Technical Specifications Page 83 Lot 7: Weighing Scale, Digital</p> <p>xxx</p> <p>Loading weight capacity: at least 500lbs/200kg x 0.2 lbs display for weight and <u>BMI</u></p> <p>xxx</p>	<p><i>As clarified by the End-User</i></p>

<p>Section VII. Technical Specifications Page 85 Lot 9: pH Meter, Portable</p> <p>xxx</p> <p>Power: battery operated ≠with suitable AC adaptor for 220V/60hz</p> <p>xxx</p> <p>pH Calibration: Min 2 points automatic or manual with buffer solution for 2 min <u>calibration</u></p> <p>xxx</p>	<p><i>As clarified by the End-User</i></p>
<p>Section VII. Technical Specifications Page 88 Lot 12: Trinocular Microscope with Monitor</p> <p>Xxx</p> <p>0.7X – 4.5X eyepiece magnification <u>zoom objective</u></p> <p>xxx</p> <p>With compatible HD color camera</p> <ul style="list-style-type: none"> • <u>With Compatible SD Card 128GB</u> <p>xxx</p> <p>With compatible Touch screen LED Monitor, <u>UHD</u></p> <ul style="list-style-type: none"> ❖ Screen size: Atleast 41 <u>42</u> inches <p>xxx</p>	<p><i>To amend the Technical Specifications as clarified by End-User</i></p>

B. CLARIFICATIONS

The Department of Budget and Management - Procurement Service, Bids and Awards Committee hereby clarifies prospective bidders' concerns and queries:

ITEM	CONCERN	REFERENCE	CLARIFICATION/ RESOLUTION
1. GOLDEN BAT (FAR EAST) INC.			
1	Request to bring the End-User/Technical Working Group to the companies that they already sold the post qualified items, instead of submitting demo units considering the problem with shipment as caused by current viral outbreak	<p>Section VI. Schedule of Requirements Page 73</p> <p>xxx</p> <p>The Bidder shall submitted the demo units to the Joint TWG of DOH-PS DBM at the PS-DBM Complex for the hereunder lots at the supplier's warehouse within fifteen (15) calendar days after receipt of notice of Single/ Lowest Calculated Bid from the BAC.</p> <p>xxx</p>	<i>Please refer to the abovementioned amendment.</i>
2	Request to delete the requirement for the accessories like magnifying lens, marking probe and lends holder, since it is already automatic.	<p>Section VII. Technical Specifications Page 78 Lot No. 3: Colony Counter, Automated</p> <p>xxx</p> <p>Accessories:</p> <ul style="list-style-type: none"> • Marking Probe • Magnifying Lens • Lens Holder <p>xxx</p>	<i>Please refer to the abovementioned amendment.</i>
3	Bidders clarified the 0.7x-4.5x eyepiece magnification should be zoom objective.	<p>Section VII. Technical Specifications Page 88 Lot 12: Trinocular Microscope with Monitor</p> <p>xxx</p>	<i>Please refer to the abovementioned amendment.</i>

		❖ 0.7X – 4.5X eyepiece magnification xxx	
4	Bidders asked if the resolution of the camera is 3264 x 1836 pixel HD ultra-color camera	Section VII. Technical Specifications Page 88 Lot 12: Trinocular Microscope with Monitor xxx With compatible HD color camera ❖ Resolution: 3840 x 2160 xxx	<i>As clarified by the End-User, original requirement is retained.</i>

ITEM	CONCERN	REFERENCE/CONCERN	CLARIFICATION/RESOLUTION
CONCERNS RAISED DURING PRE-BID			
1	A prospective bidder inquired if they can offered an equivalent for the ELISA-verified allergen	Section VII. Technical Specifications Page 77 Lot 2: Animal Containment Workstation • ELISA-verified allergen	The requirement should be retained as clarify by End-User
2	A prospective bidder request to consider other type of filtration module.	Section VII. Technical Specifications Page 77 Lot 2: Animal Containment Workstation • ULPA filtration module	The requirement should be retained as clarify by End-User

3	A prospective bidder asked to consider a netbook with eleven (11) inches screen display for the monitor	<p>Section VII. Technical Specifications Page 88 Lot 12: Trinocular Microscope with Monitor</p> <p>With compatible Touch Screen LED Monitor</p> <ul style="list-style-type: none"> • Screen size: At least 11 inches 	Please refer to the abovementioned amendment.
4	A prospective bidder asked the End-User if there is a preferred display, whether a touch a screen with a small size display or non-touch screen display with a big size display	<p>Section VII. Technical Specifications Page 88 Lot 12: Trinocular Microscope with Monitor</p> <p>With compatible Touch Screen LED Monitor</p> <ul style="list-style-type: none"> • Screen size: At least 11 inches 	Please refer to the abovementioned amendment.

All other portions of the Bidding Documents affected by these amendments shall be made to conform to the same.

Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.


WEBSTER M. LAUREÑANA
Chairperson, Bids and Awards Committee V

Schedule of Requirements

LOT NO.	QTY/UOM	ITEM/DESCRIPTION	DELIVERY PERIOD
1	2 units	Analytical Balance	Within SIXTY (60) Calendar Days from the effective date indicated in the Notice to Proceed
2	1 unit	Animal Containment Workstation	
3	2 units	Colony Counter, Automated	
4	14 units	External Defibrillator, Automated	
5	2 units	Biological Refrigerator	
6	2 units	Hot Plate Stirrer	
7	3 units	Weighing Scale, Digital	
8	4 units	Loop Sterilizer	
9	4 units	pH Meter, Portable	
10	4 units	Refrigerator, Household 2 door	
11	2 units	Stomacher	
12	2 units	Trinocular Microscope with Monitor	

I hereby commit to comply and delivery all the above requirements in accordance with the above stated schedule.

Name of Company

Signature Over Printed Name of
Authorized Representative

Date



LOT NO. 2 : **ANIMAL CONTAINMENT WORKSTATION**
QUANTITY : **1 unit**
APPROVED BUDGET FOR THE CONTRACT : **P 624,200.00**

AGENCY SPECIFICATIONS	BIDDER'S STATEMENT OF COMPLIANCE	REFERENCE
ITEM DESCRIPTION:	Brand and Model:	
<ul style="list-style-type: none"> • ELISA-verified allergen 		
<ul style="list-style-type: none"> • Motorized height- adjustable stand 		
<ul style="list-style-type: none"> • ULPA filtration module 		
<ul style="list-style-type: none"> • With easy to lift large tray 		
<ul style="list-style-type: none"> • Dimensions: <ul style="list-style-type: none"> ❖ Width: 1219-1524mm ❖ Depth: 800-820mm ❖ Height: 1371-1676mm 		
<ul style="list-style-type: none"> • 220V power supply 		
<ul style="list-style-type: none"> • Accessories: <ul style="list-style-type: none"> • Chair, stool type, heavy duty, adjustable height, stainless steel with caster wheel (lockable). • Compatible Automatic Voltage Regulator. Standard Requirements: <ul style="list-style-type: none"> • Printing of the DOH & BOQ logo in a white vinyl sticker, 10"W x 8"H to be placed in the front right uppermost corner. • NSF Standard. 		

<ul style="list-style-type: none"> • Training Requirements: Technical Training of 2-5 end-users 		
---	--	--

I hereby certify that the statement of compliance to the foregoing technical specifications are true and correct, otherwise, if found to be false either during bid evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.

Name of Company	Signature Over Printed Name of Authorized Representative	Date
-----------------	--	------



LOT NO. 3 : **COLONY COUNTER, AUTOMATED**
QUANTITY : **2 units**
APPROVED BUDGET FOR THE CONTRACT : **P 604, 304.00**

AGENCY SPECIFICATIONS	BIDDER'S STATEMENT OF COMPLIANCE	REFERENCE
ITEM DESCRIPTION:	Brand and Model:	
<ul style="list-style-type: none"> • Darkfield colony counter 		
Display: 6 digits, 7-Seg. LED		
Counting time: at least 1,000 colonies per second		
Magnification: x 1.7 surface magnification		
Power: 230 V±10%, 60 Hz		
Dimensions: Manufacturer's standard		
Weight: Manufacturer's standard		
Accessories:		
<ul style="list-style-type: none"> • Dust Cover 		
<ul style="list-style-type: none"> • Minimum size of detected colony: 0.1mm 		
<ul style="list-style-type: none"> • Compatible for all Petri dishes: 55, 65, 90, 150mm, petrifilm, and Compact Dry. 		
Standard Requirements: <ul style="list-style-type: none"> • Printing of the DOH & BOQ logo in a white vinyl sticker, 2"L & 2"W to be placed in any area of the equipment but will not affect its function or cover any marking. 		

Training Requirements: Technical Training of 2-5 end-users		
---	--	--

I hereby certify that the statement of compliance to the foregoing technical specifications are true and correct, otherwise, if found to be false either during bid evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.

Name of Company	Signature Over Printed Name of Authorized Representative	Date
-----------------	---	------



LOT NO. 7 : **WEIGHING SCALE, DIGITAL**
QUANTITY : **3 units**
APPROVED BUDGET FOR THE CONTRACT : **P 105,000.00**

AGENCY SPECIFICATIONS	BIDDER'S STATEMENT OF COMPLIANCE	REFERENCE
ITEM DESCRIPTION:	Brand and Model:	
<ul style="list-style-type: none"> • Dimension: Manufacturer's standard 		
<ul style="list-style-type: none"> • Non sliding surface 		
Battery operated, include 1 set of compatible batteries		
Loading weight capacity: at least 500 lbs/220 kg x 0.2 lbs display for weight and BMI		
<ul style="list-style-type: none"> • Folding telescopic height meter, at least 60 to 200 cm 		
Platform dimensions: Manufacturer's standard		
Dimension: Manufacturer's standard		
Accessories: Manual, cover		
Standard Requirements: <ul style="list-style-type: none"> • Printing of the DOH & BOQ logo in a white vinyl sticker, 2"L & 2"W to be placed in any area of the equipment but will not affect its function or cover any marking. 		
Training Requirements: Training of 2-5 end-users		

I hereby certify that the statement of compliance to the foregoing technical specifications are true and correct, otherwise, if found to be false either during bid evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.

Name of Company _____ Signature Over Printed Name of _____ Date _____
 Authorized Representative

LOT NO. 9 : **pH METER, PORTABLE**
QUANTITY : **4 units**
APPROVED BUDGET FOR THE CONTRACT : **P 110,000.00**

AGENCY SPECIFICATIONS	BIDDER'S STATEMENT OF COMPLIANCE	REFERENCE
ITEM DESCRIPTION:	Brand and Model:	
Measurement type: using pH electrode with BNC end cable		
Range: Ph -2.0 to 16		
Accuracy: ±0.02 pH		
Resolution: 0.01 pH		
Operating temperature: 0-4.5°C		
Temperature compensation: Automatic or manual		
pH Calibration: Min 2 points automatic or manual with buffer solution for 2 min calibration		
Power: battery operated with suitable AC adaptor for 220V/60 hz		
Carrying case: Hard carrying case		
Display: Backlite LCD		
Standard Requirements: <ul style="list-style-type: none"> • Printing of the DOH & BOQ logo in a white vinyl sticker, 2"L & 2"W to be placed in any area of the equipment but will not affect its function or cover any marking. 		

Training Requirements: Training of 2-5 end-users		
---	--	--

I hereby certify that the statement of compliance to the foregoing technical specifications are true and correct, otherwise, if found to be false either during bid evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.

Name of Company	Signature Over Printed Name of Authorized Representative	Date
-----------------	--	------



LOT NO. 12 : **TRINOCULAR MICROSCOPE WITH MONITOR**

QUANTITY : **2 units**

APPROVED BUDGET FOR THE CONTRACT : **P 400,000.00**

AGENCY SPECIFICATIONS	BIDDER'S STATEMENT OF COMPLIANCE	REFERENCE
ITEM DESCRIPTION:	Brand and Model:	
Brand and Model		
<ul style="list-style-type: none"> • LED light 		
<ul style="list-style-type: none"> • 10X – 20X lens magnification 		
<ul style="list-style-type: none"> • 0.7X – 4.5X eyepiece magnification 		
With compatible HD color camera <ul style="list-style-type: none"> • Resolution: 3840x2160 • Video output: HDMI, USB • With SD card slot • With Compatible SD Card 128GB 		
With compatible LED Monitor <ul style="list-style-type: none"> • Screen size: At least 42 inches 		
Standard Requirements: <ul style="list-style-type: none"> • Printing of the DOH & BOQ logo in a white vinyl sticker, 2”L & 2”W to be placed at the base. 		
Training Requirements: Training of 2-5 end-users		

I hereby certify that the statement of compliance to the foregoing technical specifications are true and correct, otherwise, if found to be false either during bid evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.

Name of Company
Signature Over Printed Name of Authorized Representative
Date

